

Modes of Failure in Revision Hip and Knee Replacement




Kevin J. Bozic, MD, MBA
Harry E. Rubash, MD



J. Berry, MD

“It’s like déjà vu all over again!”



Background

- Total joint replacement is one of the most commonly performed and successful operations in orthopaedics as defined by *clinical outcomes* and *implant survivorship**



*

TWENTY-FIVE-YEAR SURVIVORSHIP OF TWO THOUSAND CONSECUTIVE PRIMARY CHARNLEY TOTAL HIP REPLACEMENTS


FACTORS AFFECTING SURVIVORSHIP OF ACETABULAR AND FEMORAL COMPONENTS

BY DANIEL J. BERRY, MD, W. SCOTT HARMSSEN, MS, MIGUEL E. CABANELA, MD, AND BERNARD E. MORREY, MD

Investigation performed at the Department of Orthopedic Surgery, Mayo Clinic, Rochester, Minnesota

Background

- Total joint replacement (TJR) is one of the most cost-effective procedures in all of medicine.



TJA Volume Estimates

Primary and Revision TJA Procedures Performed in the US






Chart 7: U.S. Population Growth (1995-2010E)



Source: Merrill Lynch, U.S. Census Bureau




U.S. TJR Payer Mix




Source: AHRQ, HCUPnet, 2002 Nationwide Inpatient Sample, <http://hcup.ahrq.gov/HCUPnet.asp>, site accessed on July 26, 2004. Total Hip Replacement is sum of ICD9-CM Procedure Codes 81.51 and 81.53. 81.51 Total Hip Replacement, 81.53 Revise Hip Replacement. NIS data is collected for calendar years (January – December). Routine discharge is discharge to home only. Discharge to another institution includes discharge to SNF and IRF.

DRG 209/471: 1998-2002



% of Medicare Discharges



% of Medicare Inpatient Charges


TJR Failure

- Despite the success achieved with most primary TJR procedures, factors related to implant longevity and a younger, more active patient population have led to a steady increase in the number of failed TJR's



Problem with Current ICD-9-CM Diagnosis Codes

- Currently, all failed TJR's are coded as either:
 - **996.4** *Mechanical complication of an internal orthopedic device, implant, or graft:*
 - Mechanical complications involving external fixation device using internal screw(s), pin(s), or other methods of fixation; grafts of bone, cartilage, muscle, or tendon; internal fixation device such as nail, plate, rod, etc.
 - **996.6** *Infection and inflammatory reaction due to internal joint prosthesis*




Problem with Current ICD-9-CM Diagnosis Codes

- New technologies and surgical techniques are constantly being introduced into the marketplace
- Despite careful laboratory testing, a certain percentage of new technologies are associated with higher rates of clinical failure
- Current ICD-9-CM Diagnosis codes limit our ability to track clinical outcomes and complications related to new techniques and technologies in TJR



TJA: Indications



Degenerative
osteoarthritis

Cartilage worn
away

Bone spurs

©MMG 2001




Arthritis—Background

- Arthritis is the second most common chronic condition in the US (sinusitis is first)
 - Most common among elderly
- 20-30% of people over age 70 suffer from osteoarthritis (OA) of the hip
- Arthritis affects over 32 million people in the US
- Total costs associated with arthritis are over \$82B/year, including hospital and drug costs, nursing home costs, and lost productivity and work

Treatment Options: Non-operative



- Activity Modification
- Weight Loss
- Cane/walker
- Physical Therapy
- Medications:
 - NSAIDs
 - COX-2 Inhibitors
 - Nutritional supplements
- Injections:
 - Corticosteroid
 - Viscosupplementation



Surgical Treatment Options


■ Joint preserving operations

- Arthroscopy
- Cartilage transplantation
- Osteotomy



■ Arthroplasty Options:

- Hemiarthroplasty
- Resurfacing arthroplasty
- **Total joint arthroplasty**





Goals of Joint Replacement Surgery


- **Relieve pain!!!**
- Restore function, mobility



Anatomy—Hip




THA Implants



Implant Choice

Cemented:


- Elderly (>65)
- Low demand
- Better early fixation
- ? late loosening



Implant Choice


Cementless:

- Younger
- More active
- Protected weight-bearing first 6 weeks
- ? Better long-term fixation





Technique: Total Hip Replacement

- Femoral neck resection




Technique: Total Hip Replacement

- Acetabular reaming
- Insertion of acetabular component




Technique: Total Hip Replacement

- Reaming/broaching of femoral component




- Insertion of femoral component



Technique: Total Hip Replacement


- Femoral head impaction

Femoral Stem
(inserted into femoral canal)





- Final implant

Artificial Hip
(in place)




Anatomy—Knee



Normal Knee Anatomy


Knee Replacement—Implants





Patellar
component





Knee Replacement—Bone Cuts



Knee Replacement—Implants



Knee Replacement—Implants




Causes of TJR Failure

- Wear of articular bearing surface
- Aseptic/mechanical loosening
- Osteolysis
- Infection
- Instability
- Peri-prosthetic fracture
- Implant Failure

Timing of TJR Failure

- Early (<10%)
 - *Dislocation*
 - Infection
 - Implant failure
- Late (> 5 yrs post op)
 - *Wear of articular bearing surface*
 - Osteolysis
 - Mechanical loosening
 - Peri-prosthetic fracture





Dislocation/Instability






Infection




Wear of Articular Bearing Surface




Osteolysis



Aseptic/Mechanical Loosening




Peri-Prosthetic Fracture



Implant Failure



Major Osseous Defects



Major Osseous Defects




Benefits of Revised Codes

MEDPAR database

- Robust source of data for evaluating clinical outcomes, complication rates, and resource utilization in TJR
- However, current ICD-9 codes do not distinguish between the type of orthopedic device failure or the cause of TJR failure

ASSOCIATION BETWEEN HOSPITAL AND SURGEON PROCEDURE VOLUME AND OUTCOMES OF TOTAL HIP REPLACEMENT IN THE UNITED STATES MEDICARE POPULATION*

BY JEFFREY N. KATZ, MD, MS, ELENA LOSINA, PhD, JANE BARRETT, MSc,
CHARLOTTE B. PHILLIPS, RN, MPH, NIZAR N. MAHOMED, MD, ScD, ROBERT A. LEW, PhD,
EDWARD GUADAGNOLI, PhD, WILLIAM H. HARRIS, MD, ROBERT POSS, MD, AND JOHN A. BARON, MD, MPH



Benefits of Revised Codes

- Ability to specify the cause of implant failure
- Ability to evaluate implant-specific TJR failure rates => refine indications, surgical technique, and implant choice



- Facilitates steady, continuous quality improvement by shortening the time span for detection of poor performance of new techniques and technologies

Benefits of Revised Codes



NATIONAL JOINT REPLACEMENT
REGISTRIES: HAS THE TIME COME?

By WILLIAM J. MALONEY, MD



American Joint Replacement Registry(AJRR)

■ Goals

- Accurately define the epidemiology of TJR in the U.S.
- Identify risk factors for poor outcomes
- To improve outcomes through continuous feedback to participating centers and surgeons

- ***The success of this project is critically dependent on having revised ICD-9-CM Codes that differentiate between different modes of failure in TJA!!***

Benefits of Revised Codes


THE SWEDISH TOTAL HIP REPLACEMENT REGISTER

By HENRIK MALCHAU, MD, PhD, PETER HERBERTS, MD, PhD,
THOMAS EISLER, MD, GÖRAN GARELLICK, MD, PhD, AND PETER SÖDERMAN, MD, PhD

- Credited with substantially reducing revision rates through early identification of failures
- Revision rate of 8% (vs. 17% in U.S.)
- Estimated that over 11,000 revisions have been avoided
 - Direct cost savings of \$140 million

Summary

- Hip and knee replacement are commonly performed and highly successful operations
- Most TJR's last 10-15 years or more





Summary

- When failure does occur, the type and cause of failure will determine the type of revision joint replacement procedure performed (partial vs. total)



Summary

- Current ICD-9-CM Diagnosis codes do not provide any information regarding the type or cause of implant failure
- Revised codes will benefit patients, providers, and payors by facilitating continuous feedback and improvement in clinical outcomes in TJR



***Thank
You!!!***



Questions??